



Emission Check

You're on the Road to Cleaner Air.

STATE OF WASHINGTON • DEPARTMENT OF ECOLOGY

Northwest Regional Office (425) 649-7000 / Eastern Regional Field Office (509) 329-3400

Southwest Regional Office (360) 407-6300 / Vancouver Field Office (360) 690-7171

TTY 711 or 1-800-833-6388

Monthly Analyzer Accuracy Report

(To be conducted every 30 days and/or after relocation or repair)

(1) Date	(2) Leak Check * P/F	(3) Bottle Pressure 25lbs +	(4) HC Value (ppm)	(5) P.E.F. (Bench)	(6) HC Target	(7) HC at probe	(8) HC Error +/- 30ppm	(9) CO% Target	(10) CO% at probe	(11) CO% Error +/- .20%	(12) Analyzer * P/F	(13) Name or initials of AES

- (If your analyzer fails leak check, HC or CO readings, you must notify your Ecology Representative within 24 hours)